

TELE-TRAFFIC (UK) LIMITED CALIBRATION DATA SHEET

LTI 20-20 TS/M SpeedScope Speed Measuring Device

Certificate No 17- 0705

MANUFACTURED BY: LASER TECHNOLOGY INC.

MANUFACTURERS SERIAL NUMBER: 014395

UNITED KINGDOM SERIAL NUMBER: 448262

Visual Inspection:	Pass
Initialisation Test:	Pass
Display Test:	Pass
Serial Port Check:	Pass

Distance Checks

No:1	25 Metres	Pass	Speed: 0 mph
No:2	75 Metres	Pass	Speed: 0 mph
No:3	100 Metres	Pass	Speed: 0 mph

Pulse Repetition Frequency Test	125 pps nominal
Output Period Specification:	7.9968 ms to 8.000 ms
Measured Output Period:	7.9984 ms

Speed Simulation Tests

Specification:	+/- 1MPH or 2 KPH	
Measured Test No:1	@ 20MPH	= 19
Measured Test No 2	@ 30MPH	= 30
Measured Test No 3	@ 50MPH	= 50
Measured Test No:4	@ 70MPH	= 69
Measured Test No:5	@ 80MPH	= 79

Scope Alignment Test:	Pass
(Horizontal and Vertical alignment correct)	

NOTES

Any parts or components that may have been replaced during the calibration procedure are certified as identical to those used in the device as type approved.

I CERTIFY THAT THE FOREGOING INTERNAL TESTS AND TEST DATA ARE CORRECT AND THAT THE INSTRUMENT ALIGNMENT IS CORRECT

TESTED BY: James Sample
[Signature]

DATE: : 11 August 2017
CAL DUE: : 11 August 2018

TIME: 10:07 AM

TELE-TRAFFIC (UK) LIMITED

LaserTec Centre
C2 Harris Road
WARWICK
CV34 5JU

Tel: 01926-407272
Fax: 01926-407977
Email: tt @ teletraffic . co. uk

CERTIFICATE OF CONFORMITY

Certificate Number: **17-0705**

Date: **11 August 2017**

Customer: **STAFFORDSHIRE POLICE**

Instrument : **LTI 20-20 TS/M SpeedScope Speed Measuring Device**

Manufacturer: **LASER Technology inc**

Serial Number: **014395**

U.K. Serial Number: **448262**

This instrument has been calibrated/calibration verified and complies with the manufacturers specification at the measured points. Calibration/verification was controlled by the use of documented procedures using equipment traceable to National standards.

Approved Signatory



for Tele-Traffic (UK) Ltd.