

CMAG  
TRUCAM (2)

# TELE-TRAFFIC (UK) LIMITED CALIBRATION DATA SHEET

LTI 20-20 TruCAM

Certificate No 18- 0295

MANUFACTURED BY: LASER TECHNOLOGY INC.

MANUFACTURERS SERIAL NUMBER: TC 005996

UNITED KINGDOM SERIAL NUMBER: 449666TC

|                      |      |
|----------------------|------|
| Visual Inspection:   | Pass |
| Initialisation Test: | Pass |
| Display Test:        | Pass |
| Serial Port Check:   | Pass |

#### Distance Checks

|      |            |      |              |
|------|------------|------|--------------|
| No:1 | 25 Metres  | Pass | Speed: 0 mph |
| No:2 | 75 Metres  | Pass | Speed: 0 mph |
| No:3 | 100 Metres | Pass | Speed: 0 mph |

|                                  |                       |
|----------------------------------|-----------------------|
| Pulse Repetition Frequency Test: | 200.321 ppm nominal   |
| Output Period Specification:     | 200.221 to 200.421 Hz |
| Measured Output Period:          | 200.321 Hz            |

#### Speed Simulation Tests

|                    |                   |      |
|--------------------|-------------------|------|
| Specification:     | +/- 1MPH or 2 KPH |      |
| Measured Test No:1 | @ 20MPH           | = 19 |
| Measured Test No 2 | @ 30MPH           | = 29 |
| Measured Test No 3 | @ 50MPH           | = 50 |
| Measured Test No:4 | @ 70MPH           | = 69 |
| Measured Test No:5 | @ 80MPH           | = 80 |

#### Scope Alignment Test:

(Horizontal and Vertical alignment correct)

Pass


#### NOTES

Any parts or components that may have been replaced during the calibration procedure are certified as identical to those used in the device as type approved.

I CERTIFY THAT THE FOREGOING INTERNAL TESTS AND TEST DATA ARE CORRECT  
AND THAT THE INSTRUMENT ALIGNMENT IS CORRECT

TESTED BY:

James Sample



DATE: : 26 March 2018  
CAL DUE: : 26 March 2019

TIME: 3:16 PM

## TELE-TRAFFIC (UK) LIMITED

LaserTec Centre  
C2 Harris Road  
WARWICK  
CV34 5JU

Tel: 01926-407272  
Fax: 01926-407977  
Email: tt @ teletraffic . co. uk

### CERTIFICATE OF CONFORMITY

Certificate Number: 18 -0295

Date: 26 March 2018

Customer: CENTRAL MOTORWAY POLICE GROUP

Instrument : LTI 20-20 TruCAM

Manufacturer: LASER Technology inc

Serial Number: TC 005996

U.K. Serial Number: 449666TC

This instrument has been calibrated/calibration verified and complies with the manufacturers specification at the measured points. Calibration/verification was controlled by the use of documented procedures using equipment traceable to National standards.

Approved Signatory



for Tele-Traffic (UK) Ltd.