

# TELE-TRAFFIC (UK) LIMITED

LaserTec Centre  
C2 Harris Road  
WARWICK  
CV34 5JU

Tel: 01926-407272  
Fax: 01926-407977  
Email: tt @ teletraffic . co. uk

## CERTIFICATE OF CONFORMITY

Certificate Number: 19 -0380

Date: 10 April 2019

Customer: CENTRAL MOTORWAY POLICE GROUP

Instrument : LTI 20-20 TruCAM


Manufacturer: LASER Technology inc

Serial Number: TC 005996

U.K. Serial Number: 449666TC

This instrument has been calibrated/calibration verified and complies with the manufacturers specification at the measured points. Calibration/verification was controlled by the use of documented procedures using equipment traceable to National standards.

Approved Signatory



for Tele-Traffic (UK) Ltd.

# TELE-TRAFFIC (UK) LIMITED

## CALIBRATION DATA SHEET

LTI 20-20 TruCAM

Certificate No 19- 0380

MANUFACTURED BY: LASER TECHNOLOGY INC.

MANUFACTURERS SERIAL NUMBER: TC 005996

UNITED KINGDOM SERIAL NUMBER: 449666TC

Visual Inspection: Pass  
Initialisation Test: Pass  
Display Test: Pass  
Firmware Version Check: Pass

### Distance Checks

No:1	25 Metres	Pass	Speed: 0 mph
No:2	75 Metres	Pass	Speed: 0 mph
No:3	100 Metres	Pass	Speed: 0 mph

**Pulse Repetition Frequency Test** 200.321 ppm nominal  
Output Period Specification: 200.221 to 200.421 Hz  
Measured Output Period: 200.321 Hz

### Speed Simulation Tests

Specification: +/- 1MPH or 2 KPH

Measured Test No:1	@ 20MPH	= 19
Measured Test No 2	@ 30MPH	= 30
Measured Test No 3	@ 50MPH	= 49
Measured Test No:4	@ 70MPH	= 69
Measured Test No:5	@ 80MPH	= 80

**Scope Alignment Test:** Pass  
(Horizontal and Vertical alignment correct)

### NOTES

Any parts or components that may have been replaced during the calibration procedure are certified as identical to those used in the device as type approved.

**I CERTIFY THAT THE FOREGOING INTERNAL TESTS AND TEST DATA ARE CORRECT AND THAT THE INSTRUMENT ALIGNMENT IS CORRECT**

TESTED BY: Paul Garratt

DATE: : 10 April 2019  
CAL DUE: : 10 April 2020

TIME: 12:12 PM